

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in Article 13 of Directive 2014/31/EU, after
having established that the measuring instrument meets the applicable
requirements of Directive 2014/31/EU, to:

Manufacturer Precisa Gravimetrics AG
Moosmattstrasse 32
CH-8953 Dietikon
Switzerland

Measuring instrument **A Non-automatic weighing instrument**
Type : Serie LX/LS/LT

Further properties are described in the annexes:

- Description T7911 revision 3;
- Documentation folder T7911-4.

Valid until 13 June 2023

Remarks This revision replaces the earlier versions, including its documentation
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
28 November 2017



C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at
[http://ec.europa.eu/growth/tools-
databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, shall not be in conflict with the legislation.

1.1 Essential parts

The electronics;
 The mechanical assembly with load cell.

See block diagram:

Number	Pages	Description	Remarks
7911/0-07	1	Block diagram	-

EMI protection measures:

- The instrument has a metal enclosure.

1.2 Essential characteristics

Accuracy class	(I)		(II)	
Maximum capacity	$120 \text{ g} \leq \text{Max} \leq 1220 \text{ g}$		$160 \text{ g} \leq \text{Max} \leq 10200 \text{ g}$	
Verification scale interval	$e \geq 1 \text{ mg}$		$e \geq 10 \text{ mg}$	
Weighing range(s)	Single interval			
Maximum number of scale intervals	$n \leq 320000$ divisions			
Tare	$T \leq -\text{Max}$			
Temperature range	$+15 \text{ }^\circ\text{C} / +25 \text{ }^\circ\text{C}$		$+10 \text{ }^\circ\text{C} / +30 \text{ }^\circ\text{C}$	
Power supply voltage	100 – 240 V AC 50/60 Hz (with external adapter) 12 V DC (supplied by external adapter)			
Software identification	Type LX/LS	Version number: xx.xx ExxD01		$x = 0, 1, \dots, Z$
	Type LT	Version number: xx.xx ExxD00		

The software identification is displayed at start-up;
 The non-automatic weighing instrument has embedded software;

1.3 Essential shapes

Number	Pages	Description	Remarks
7911/1-01	5	Outline drawings	LS-LX-LT
7911/3-01	3	Outline drawings (alternative)	LS-LX-LT
7911/1-02	6	Exploded views	LS-LX-LT
7911/2-01	5	Exploded views (alternative)	LS-LX-LT

The data plate is secured against removal by sealing or will be destroyed when removed.

Inside the cabinet is an adjustment lock, located on the main board.

1.4 Conditional parts

The non-automatic weighing instrument may be equipped with peripheral equipment which is used for the applications listed in Article 1(2), (a) to (f) of Directive 2014/31/EU, provided that the peripheral equipment is certified to be connected to a non-automatic weighing instrument by a Notified Body responsible for type examination under Directive 2014/31/EU, or, that the equipment and the use of the equipment complies with the requirements of WELMEC 2.5 Issue 2 clause 2.2.

The non-automatic weighing instrument is fitted with a levelling device and a level indicator, unless the instrument is installed in a fixed position. A ring on the level indicator indicates when the maximum tilt is exceeded.

1.5 Non-essential parts

The non-automatic weighing instrument may be connected to non-essential devices, for example but not limited to bar code readers, foot switches, second displays and cash drawers, provided that:

- They do not present primary data used for purposes mentioned in Article 1(2), (a) to (f) of Directive 2014/31/EU unless the "Preliminary observation" in Annex I of the Directive is satisfied;
- They do not lead to an instrument having other essential characteristics than those fixed by this certificate.

Other non-essential parts:

- AC/DC adapter;
- Draft shield.

2 Information about the main constituent parts of the non-automatic weighing instrument

2.1 The electronics

2.1.1 Essential parts

Number	Pages	Description	Remarks
7911/1-03	9	Main board parts list & layout	-

2.1.2 Essential characteristics

List of legally relevant functions:

- Determination stability of equilibrium;
- Indication of stable equilibrium;
- Zero indicator;
- Semi-automatic zero-setting;
- Initial zero-setting;
- Zero-tracking;
- Semi-automatic subtractive tare balancing;
- Preset tare;
- Adjustment / set-up mode via a switch on the main board;
- Automatic span adjustment with internal adjustment mass;
operational when:
 - After switch on
 - $\Delta t \geq 3 \text{ }^\circ\text{C}$
 - On every ≤ 24 hours
- Semi-automatic span adjustment with internal adjustment mass;
- Semi-automatic span adjustment with external adjustment mass;
- Acting upon significant faults;
- Checking the display;
- Check weighing mode;
- Weighing unstable samples;
- Weight unit selection (g, mg, ct).

When equipped with a printer the following legally relevant functions may be present:

- Indications other than primary indications;
- Indication of additional information;
- Memory storage.

2.1.3 Conditional parts

Number	Pages	Description	Remarks
7911/0-04	2	Interface board layout & parts list	-

The non-automatic weighing instrument may be equipped with one or more of the following protective interfaces that have not to be secured:

- RS232;
- USB.

2.1.4 Non-essential parts

Display;
 Keyboard;
 Printer.

2.2 The mechanical assembly with load cell

2.2.1 Essential parts

Number	Pages	Description	Remarks
7911/0-06	5	Weighing cell assembly	-
7911/2-02	5	Weighing cell assembly (alternative)	-

2.2.2 Essential characteristics

Number	Pages	Description	Remarks
7911/0-08	2	Weighing cell specifications	-

2.2.3 Essential shapes

See 2.2.1.

3 Seals

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the drawing:

Number	Pages	Description	Remarks
7911/0-05	1	Sealing	-

4 Conditions for conformity assessment

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfil the requirements of point 1 of Annex III of Directive 2014/31/EU.