

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2009/23/EC, after
having established that the measuring instrument meets the applicable
requirements of Directive 2009/23/EC, to:

Manufacturer Precisa Gravimetrics AG
Moosmattstrasse 32
CH-8953 Dietikon
Switzerland

Measuring instrument **A Non-automatic weighing instrument**
Type : Serie LX/LS

Further properties are described in the annexes:
– Description T7911 revision 0;
– Documentation folder T7911-1.

Valid until 13 June 2023

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
13 June 2013

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMI (see www.nmi.nl).

Reproduction of the complete
document only is permitted.

1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, shall not be in conflict with the legislation.

1.1 Essential parts

See drawing "Block diagram", drawing number 7911/0-07;
 The electronics;
 The mechanical assembly with load cell.

EMC protection measures:

- The instrument has a metal enclosure.

1.2 Essential characteristics

Accuracy class	I	II
Maximum capacity	120 g ≤ Max ≤ 1220 g	160 g ≤ Max ≤ 10200 g
Verification scale interval	e ≥ 1 mg	e ≥ 10 mg
Maximum number of scale intervals	n ≤ 320000 divisions (per partial weighing range)	
Maximum partial weighing ranges	1	
Temperature range	+15 °C / +25 °C	+10 °C / +30 °C
Tare	T ≤ -Max	
Weighing range(s)	Single interval	
Power supply voltage	100 – 240 V AC 50/60 Hz (with external adapter) 12 V DC (supplied by external adapter)	
Software identification	Version number: xx.xx ExxD01 where xx can be 00-99 or A-Z, represents the non legally relevant parts. E..D01. represent the legal-relevant part of the software	

Software:

- The identification number will be displayed at start-up.



Description

Number **T7911** revision 0
Project number 11200389
Page 2 of 4

1.3 Essential shapes

The non-automatic weighing instrument is built according to the drawings:

- "Outline drawings", drawing number 7911/0-01;
- "Exploded view", drawing number 7911/0-02.

The descriptive markings plate is secured against removal by sealing or will be destroyed when removed.

Inside the cabinet is a calibration lock, located on the main board.

1.4 Conditional parts

The non-automatic weighing instrument may be equipped with peripheral equipment which is used for the applications listed in article 1(2)(a) of Directive 2009/23/EC, provided that the peripheral equipment is certified to be connected to an EC type-approved non-automatic weighing instrument by a Notified Body responsible for type examination under Directive 2009/23/EC, or that the equipment and the use of the equipment comply with the requirements of WELMEC 2.5 Issue 2 Section 2.2.

The non-automatic weighing instrument is fitted with a levelling device and a level indicator, unless the instrument is installed in a fixed position. A ring on the level indicator indicates when the maximum tilt is exceeded.

1.5 Non-essential parts

The non-automatic weighing instrument may be connected to non-essential devices, for example but not limited to bar code readers, foot switches, second display's and cash drawers, provided that:

- They do not present primary data used for purposes mentioned in article 1(2)(a) of Directive 2009/23/EC unless the "preliminary observations" in Annex 1 of this directive is satisfied;
- They do not lead to an instrument having other essential characteristics than those fixed by this type-approval document.

Other non-essential parts:

- AC/DC-adapter.

2 Information about the main constituent parts of the non-automatic weighing instrument

2.1 The electronics

2.1.1 Essential parts

Number	Pages	Description	Remarks
7911/0-03	5	Main board	PCB and parts list

2.1.2 Essential characteristics

List of legally relevant functions:

- Determination stability of equilibrium;
- Indication of stable equilibrium;
- Zero indicator;
- Semi-automatic zero-setting;
- Initial zero-setting;
- Zero-tracking;
- Semi-automatic subtractive tare balancing;
- Preset tare;
- Calibration / set-up mode via a switch on the main board;
- Automatic span adjustment with internal calibration mass;
operational when:
 - After switch on
 - $\Delta t \leq 3 \text{ }^\circ\text{C}$
 - On every ≤ 24 hours
- Semi-automatic span adjustment with internal calibration mass;
- Semi-automatic span adjustment with external calibration mass;
- Acting upon significant faults;
- Checking the display;
- Check weighing mode;
- Weighing unstable samples;
- Weight unit selection (g, mg, ct);

When equipped with a printer the following legally relevant function may be present:

- Indications other than primary indications;
- Indication of additional information;
- Memory storage;
- Non-weighed articles;
- Totalization;
- Multi-vendor;
- Price labeling.

2.1.3 Conditional parts

Number	Pages	Description	Remarks
7911/0-04	2	Interface board	PCB and parts list

The non-automatic weighing instrument may be equipped with one or more of the following protective interfaces that have not to be secured:

- RS232;
- USB.

2.1.4 Non-essential parts

Display;
 Keyboard;
 Printer.

2.2 The mechanical assembly with load cell

2.2.1 Essential parts

Number	Pages	Description	Remarks
7911/0-06	5	Weighing cell assembly	-

2.2.2 Essential characteristics

See document "Weighing cell specifications", drawing number 7911/0-08.

2.2.3 Essential shapes

See drawings:

- "Weighing cell assembly", drawing number 7911/0-06.

3 Seals

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the drawing:

- "Sealing", drawing number 7911/0-05.

4 Conditions for conformity assessment

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfill the requirements of article 1 of Annex IV of Directive 2009/23/EC.