Precisa

■ The Balance of Quality ■

TESTCERTIFICATE

model: **IBK 60000G**
serial number: 5910-43

---

max. weight: 62000 g

d: 1 g
e: 10 g

approved: ☐ yes ☒ no
int. calibration: ☒ yes ☐ no
accumulator: ☐ yes ☒ no

☒ single range- ☐ dual range- ☐ floating range balance

---

corner load check nominal value = 18000 g	limit: 2 g
centre	**0 g**
front left	**1 g**
back left	**1 g**
back right	**1 g**
front right	**0 g**

measurements performed automatically

calibration executed with weight set of	
calibration weight class	**F1**
serial number	**8541**
approved until	**2008-5**

reproducing check series of 10 measurements	limit: 1.0 g
nominal value	60000 g
standard deviation	**0.0 g**

measurements performed automatically

linearity check check value	limit: 2 g
0 g	**0 g**
6000 g	**0 g**
12000 g	**0 g**
18000 g	**0 g**
24000 g	**0 g**
30000 g	**0 g**
36000 g	**0 g**
42000 g	**0 g**
48000 g	**0 g**
54000 g	**0 g**
60000 g	**0 g**

measurements performed automatically

Date: 2007-09-11

Signature: S. Talarico

Controlled: S. Wander

Precisa Gravimetrics AG
Moosmattstrasse 32
Postfach 352
CH-8953 Dietikon

Precisa Gravimetrics AG Moosmattstrasse 32 Postfach 352 CH-8953 Dietikon
Phone +41 44 744 28 28 Fax +41 44 744 28 38 www.precisa.com service@precisa.ch